

Date: 14/10/2019

### **Addendum 03**

With reference to the tender notification for “**Raman spectrometer**”, **Tender Notification No.: NITK/CRF/AS/Raman/04, Dated: 28/08/2019**, Please note the following corrections:

Under “**Tender Document**”, “**Addendum 2 (dated: 04/10/2019)**”, Under “**Queries & Reply/Clarification**”, S. No. 3, For the clause “XYZ stage travel range – 100mmx100mmx25mm Step size: 100nm in XY and 20nm in Z”, Under the column “NITK’s response to the request”.

Committee decided to modify it as ‘112mmx76mmx25mm 50nm in XY and 8nm in Z’ may be read as “Committee decided to modify it as “**XYZ stage travel range: 50mmx50mmx25mm or higher. Step size: 100nm or better in XY and 8nm or better in Z**”

Last date for request tender document : 11/11/2019, before 3.00 p. m.  
Last date for Bid submission : 11/11/2019, before 4.00 p. m.  
Bid opening date(tentative) : 18/11/2019 @ 3.00 p.m.

Sd/-  
Buyer  
(Prof. Anandhan Srinivasan)

Sd/-  
Chairman  
Central Research Facility  
NITK, Surathkal