

Addendum 01

The Pre – Bid Meeting was held on **05/03/2020** at **4.30 p.m.** in the **Board Room**, NITK Surathkal for the purchase of “**Spectroradiometer**” (**Tender Notification No: NITK/CRF/Spectro/04 Dated: 17/02/2020**). The following queries were discussed & the Reply/Clarification given to the prospective bidders.

Queries & Reply/Clarification

Sl. No.	NITK Tender Specifications	Questions asked by the vendor	Reply/Clarification	Changes to the Tender
1.	Laptop configuration: 9th Gen Gemcut 15T, 15.6" Touch 4K UHD, i7- 9750H Hexacore, NVIDIA GeForce GTX 1650 (4GB), 512GB NVMe SSD, 16GB RAM, 2 TB HDD, Win 10 Pro, B&O Speakers, USB, HDMI, SD card reader, MS Office Professional with key, Extra wireless keyboard & mouse.	A suggestion given by the vendor to opt for better laptop configuration	It has been agreed to opt for better laptop configuration	Committee decided to modify it as: “Laptop configuration: 9th Gen Gemcut 15T, 15.6" Touch 4K UHD, i7- 9750H Hexacore, NVIDIA GeForce GTX 1650 (4GB), 512GB NVMe SSD, 16GB RAM, 2 TB HDD, Win 10 Pro, B&O Speakers, USB, HDMI, SD card reader, MS Office Professional with key, Extra wireless keyboard & mouse or better ”
2.	Annexure I – Checklist for Bidders. Clause 2. License Certificate of manufacturer/ supply of the item & factory License, Registration of EPF and ESI	Clause 2. License Certificate of manufacturer/ supply of the item & factory License, Registration of EPF and ESI Since this being an imported item from UK, we hope this certificate is not applicable. Please confirm.	Authorization Certificate is required & the Dealer should be registered with EPF & ESI and a certified document should be provided.	No change in Tender Document

3.	Annexure I – Checklist for Bidders Previous Supply Order as per Annexure-'B'	This being two bid, we hope furnishing the copy of the PO with the technical bid will reveal the pricing. Hope you will consider the user list and performance certificate to comply with this. Please confirm.	In the Technical Bid, a bidder can provide the PO without price details but it should be supplied if it is called at relevant stages.	No change in Tender Document
4.	Annexure I – Checklist for Bidders Clause 11. Integrity Pact Annexure D	Clause 11. Integrity Pact Annexure D This is not applicable. Please confirm	It should be provided.	As per the Tender Document
5.	EARNEST MONEY DEPOSIT Annexure –G In case of exemption from submission of EMD, proof of registration with NSIC/MSME	We are herewith attaching our MSME certificate. Hope with this we are eligible for EMD exemption. Please confirm	As per the Govt. Rule MSME Certificate holder can exempt from EMD Please refer Tender Document, Page No: 7/32, SECTION 1 B: INSTRUCTION TO BIDDER (ITB), Point No: 3, Clause m.	No change in Tender Document
6.	Commercial	Hope the payment through LC is acceptable, since it is an imported item. Please confirm.	Payment will be made through LC after confirming Performance Security.	As per the Tender Document
7.	SECTION 1 B: INSTRUCTION TO BIDDER (ITB): Clause 10	Clause 10 – Performance Security – If we are successful, a performance security of 10% of the quoted price to be submitted as Bank Guarantee to be submitted within 30 days of purchase order. Please let me know the time period required to open the letter of credit. Otherwise, we request you to consider the performance security along with the shipment before encashing the LC. This will help us to manage the BG effectively.	Recommendation for LC will be done immediately after getting the performance security in the form of Bank Guarantee. However, Canara Bank will be making the LC for NITK.	No change in Tender Document

8.	SECTION 1 B: INSTRUCTION TO BIDDER (ITB): Clause 12	Clause 12 – The OEMs/Authorized Distributors and Dealers must attach a certificate of sales and service facilities, escalation support for on-call service or station engineer, etc. Please let us know if this certificate to be submitted with the technical Bid.	Yes, all the necessary documents should be submitted with Technical Bid	As per the Tender Document
9.	SECTION 3: SCHEDULE OF REQUIREMENTS, SPECIFICATIONS AND ALLIED DETAILS Warranty Period (in months): 36 Months (3 Years)	Section 3 - Warranty Period - 3 years (36 months) Our Principal's standard warranty is 12 months. Our Principal will extend the warranty for two more years at an additional cost. However, we suggest you to choose AMC instead of extended warranty which will be quoted at a nominal cost. Please let us know your views.	No change in Tender Document. Please refer Tender Document, Page No: 31/32, Additional Mandatory Requirements	As per the Tender Document

Following points to be included in the Addendum 01:

- Under **Tender Document**, Annexure-H, page no. 30/32, Under Detailed Technical Specifications, **Under Features, Input path configuration:**
“Permanently connected Rugged Fiberoptic cable input to channel directly the light from the target to the instrument” may be read as **“Permanently connected or interchangeable Rugged Fiberoptic cable input to channel directly the light from the target to the instrument”**
- Under Tender Document, Annexure-H, page no. 30/32, Under Detailed Technical Specifications, **Under Features, Under Field of View:**
“Fiberoptic cable with Standard 25-degree field of view. Optional narrower field of view fiber optics (4 degree and 8 degree without changing the fiberoptic cable)” may be read as **“Fiberoptic cable with Standard 25-degree field of view. Optional narrower field of view fiber optics (4 degree and 8 degree with or without changing the fiberoptic cable)”**.
- Under Tender Document, Annexure-H, page no. 30/32, Under Detailed Technical Specifications, **Under Features, Under Communication Interface:**

“10/100 Mbps fast Ethernet interface and High-speed Wireless LAN interface” may be read as “**10/100 Mbps fast Ethernet interface and High-speed Wireless LAN interface/USB**”.

- Under Tender Document, Annexure-H, page no. 30/32, Under Detailed Technical Specifications, **Under Software, Under Data Acquisition Software and preliminary data processing software:**

“Spectral acquisition software and provision to save spectrum along with GPS co-ordinates from an external GPS along with the spectrum data” may be read as “**Spectral acquisition software and provision to save spectrum along with GPS co-ordinates from an internal/external GPS along with the spectrum data**”

It is decided to extend the Bid submission date by following dates

Last date for request tender document : 17/04/2020, before 3.00 p. m.
Last date for Bid submission : 17/04/2020, before 4.00 p. m.
Bid opening date : 21/04/2020 @ 3.00 p.m.

Sd/-
Buyer
(Dr. Shwetha H. R.)

Sd/-
Chairman
Central Research Facility
NITK, Surathkal