

Corrigendum No 1

Date: 21/02/2023

This is with reference to the tender notification for “Analytical Balance (0.1 mg to 160 g)”,
NIQ notification **Reference No: Met-Mat/AS/AB/IRGCF/993/2022-23/04** **Dated: 07/02/2023**

There are following additions to specification.

1. Minimum weighing value: 1 mg.
2. Weighing Capacity: 160 g
3. Internal Calibration- Yes
4. Breeze shield – Yes

Sd/-
Dr. S Anandhan
(Buyer)

Sd/-
HOD, Dept. of MME
NITK, Surathkal