



**NATIONAL INSTITUTE OF TECHNOLOGY KARNATAKA,
SURATHKAL, SRINIVASNAGAR, MANGALORE-575025**

Date: 06/11/2023

Corrigendum No. 01

With reference to the Bid for “**Supply and Installation of 1/4” SS316 gas tubing materials**”
for Notification Number. “**NITK/ME/SERB/23/GT/04**”

Dated: 06-11-2023

please note the following:

Additions/corrections to Technical Specification:

- 1) 4 individual lines of SS316 1/4” OD Tubing should be done for Gases H₂, O₂, N₂ and Compressed Air. Total length of 150 meters to consider.
- 2) High Pressure Regulator and Line Regulator should be compatible for above mentioned Gases
- 3) Details of product offered in the bid are to be uploaded along with the bid documents. Buyer can match and verify the datasheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.
- 4) Bidder should have dedicated Service & Support centre in Karnataka for service support

Sd/-
(Dr. Poornesh Kumar Koorata)
Signature of Buyer

Sd/-
(Prof. Ravikiran Kadoli)
Signature of HOD
(Dept. of Mechanical Engineering)